

# Product Specification

Standard number : DK-010  
Issue number : BIF17  
Code number : 44006685

Name : NPW Span calibration solution No.10  
(TN:0~5mg/L TP:0~5mg/L)

(Other name : )  
Grade : E P  
Volume : 5 0 0 m L

3-2-12 UCHIHIRANOMACHI, CHUO-KU,  
OSAKA 540-0037, JAPAN

HAYASHI PURE CHEMICAL IND.,LTD.

No.	Item	(unit)	Specification	Test Procedure
1	Appearance (Form)	-	Liquid	Visual method
2	(Hue)	-	Colorless	Visual method
3	Total nitrogen	mg/L	4. 5 ~ 5. 5	Ion Chromatography
4	Total phosphorus	mg/L	4. 5 ~ 5. 5	Ion Chromatography
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				
32				
33				
34				
35				
36				
37				
38				
39				
40				

remark T/D Code 143D470