

# Product Specification

Name : NPW Span calibration solution No.40 (TN:0~200mg/L TP:0~5mg/L)

Another name :

Grade : EP

Volume : 500mL

Specification number : DK-040

Issue number : BIF17

Code number : 44006985

HAYASHI PURE CHEMICAL IND.,LTD.

3-2-12 UCHIHIRANOMACHI,CHUO-KU, OSAKA 540-0037, JAPAN

No.	Item	(unit)	Specification	Test Procedure
1	Appearance (Form)	-	Liquid	Visual method
2	Appearance (Hue)	-	Colorless	Visual method
3	Total nitrogen	mg/L	190 ~ 210	Ion chromatography
4	Total phosphorus	mg/L	4.5 ~ 5.5	Ion chromatography
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				
32				
33				
34				
35				
36				
37				
38				
39				
40				

remark

T/D Code 143D500