

Product Specification

Name : NPW Span calibration solution No.44 (TN:0~30mg/L TP:0~4mg/L)

Another name :

Grade : EP

Volume : 500mL

Specification number : DK-044

Issue number : CDD28

Code number : 44015815

HAYASHI PURE CHEMICAL IND.,LTD.

3-2-12 UCHIHIRANOMACHI,CHUO-KU, OSAKA 540-0037, JAPAN

No.	Item	(unit)	Specification	Test Procedure
1	Appearance (Form)	-	Liquid	Visual method
2	Appearance (Hue)	-	Colorless	Visual method
3	Total nitrogen	mg/L	28.5 ~ 31.5	Ion chromatography
4	Total phosphorus	mg/L	3.6 ~ 4.4	Ion chromatography
5				
6				
7				
8				
9				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				
31				
32				
33				
34				
35				
36				
37				
38				
39				
40				

remark

T/D Code 143D038